OD-33 (9-2003)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	Year
JANUARY	2011

Instrument Location: LANSING POLICE

Instrument Number	ORI Number
931024	MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
1-5-11	SIMULATOR	MI 335900	(1) .082	H0701
		Certification Number 10636	(2) —	Signature D. Dahlke #397
1-12-11	Simulator	MI 3351900	(1) .081	F0801
		Certification Number 10636	(2) —	Signature D. Dahlke #397
1-18-11	Simulator	MI 3351900	(1) .082	F0801
		Certification Number 10636	(2) —	Signature D. Dahlke #397
1-26-11	Simulator	MI 3351900	(1) .081	F0801
		Certification Number 10636	(2) —	Signature D. Dahlke #397
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature

I have inspected, serviced, and checked this instrument and certify that it is in proper working order.
Instrument Manufacturer's Authorized Representative

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature: D. Dahlke #397

Date: 2-1-11

OD-33 (9-2003)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	Year
FEBRUARY	2011

Instrument Location: LANSING POLICE

Instrument Number	ORI Number
931024	MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
2-1-11	Simulator	MI 3351900	(1) .082	F0801
		Certification Number 10636	(2) —	Signature D. Dahlke #397
2-7-11	Simulator	MI 3351900	(1) .080	F0801
		Certification Number 10636	(2) —	Signature D. Dahlke #397
2-15-11	Simulator	MI 3351900	(1) .082	H0701
		Certification Number 10636	(2) —	Signature D. Dahlke #397
2-22-11	Simulator	MI 3351900	(1) .082	H0701
		Certification Number 10636	(2) —	Signature D. Dahlke #397
2-28-11	Simulator	MI 3351900	(1) .083	H0701
		Certification Number 10636	(2) —	Signature D. Dahlke #397
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature: D. Dahlke #397

Date: 3-6-11

OD-33 (9-2003)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

Month	Year
MARCH	2011

Instrument Location: LANSING POLICE

Instrument Number: 931024

ORI Number: MI 3351900

EVIDENTIAL BREATH TESTING LOG

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
3-6-11	Simulator	MI 3351900	(1) .084	H0701
		Certification Number 10636	(2) —	Signature D. Dahlke #397
3-13-11	Simulator	MI 3351900	(1) .082	H0701
		Certification Number 10636	(2) —	Signature D. Dahlke #397
3-22-11	Simulator	MI 3351900	(1) .083	H0701
		Certification Number 10636	(2) —	Signature D. Dahlke #397
3-30-11	Simulator	MI 3351900	(1) .084	H0701
		Certification Number 10636	(2) —	Signature D. Dahlke #397
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature: D. Dahlke #397

Date: 3-30-11

OD-33 (9-2003)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	Year
APRIL	2011

Instrument Location: LANSING POLICE

Instrument Number	ORI Number
931024	MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
4-6-11	Simulator	MI 3351900	(1) .082	H0701
		Certification Number: 10636	(2) —	Signature: D. Dahlke #397
4-12-11	Simulator	MI 3351900	(1) .083	H0701
		Certification Number: 10636	(2) —	Signature: D. Dahlke #397
21	I have inspected, serviced, and checked this instrument for accuracy and certify that it is in proper working order. Marvin Gier 120 day Instrument Manufacturer's Authorized Representative	MI /	(1) .078	Ampoule Control Number: J0201 Guth-080 lot 09346 exp 12-2-11
		Certification Number: 99998	(2) —	Signature: Marvin Gier
4-28-11	Simulator	MI 3351900	(1) .082	H0701
		Certification Number: 10636	(2) —	Signature: D. Dahlke #397
		MI	(1)	
			(2)	
		MI	(1)	
			(2)	
		MI	(1)	
			(2)	
		MI	(1)	
			(2)	

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature	Date
D. Dahlke #397	5-6-11

OD-33 (9-2003)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

Month	Year
MAY	2011

Instrument Location: LANSING POLICE

Instrument Number: 931024

ORI Number: MI 3351900

EVIDENTIAL BREATH TESTING LOG

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
5-6-11	Simulator	MI 3351900	(1) .082	H0701
		Certification Number: 10636	(2) —	Signature: D. Dahlke #397
5-13-11	Simulator	MI 3351900	(1) .082	H0701
		Certification Number: 10636	(2) —	Signature: D. Dahlke #397
5-20-11	Simulator	MI 3351900	(1) .082	H0701
		Certification Number: 10636	(2) —	Signature: D. Dahlke #397
5-27-11	Simulator	MI 3351900	(1) .082	H0701
		Certification Number: 10636	(2) —	Signature: D. Dahlke #397
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature: D. Dahlke #397

Date: 5-27-11

OD-33 (9-2003)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	Year
JUNE	2011

Instrument Location: LANSING POLICE

Instrument Number: 931024

ORI Number: MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
6-4-11	Simulator	MI 3351900	(1) .084	#0701
		Certification Number: 10636	(2) —	Signature: D. Dahlke #397
6-11-11	Simulator	MI 3351900	(1) .081	G0101
		Certification Number: 10636	(2) —	Signature: D. Dahlke #397
6-23-11	Simulator	MI 3351900	(1) .081	G0101
		Certification Number: 10636	(2) —	Signature: D. Dahlke #397
6-28-11	Simulator	MI 3351900	(1) .081	G0101
		Certification Number: 10636	(2) —	Signature: D. Dahlke #397
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number: MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature: D. Dahlke #397

Date: 7-7-11

OD-33 (9-2003)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

Month	Year
JULY	2011
Instrument Location: LANSING POLICE	
Instrument Number: 931024	ORI Number: MI 3351900

EVIDENTIAL BREATH TESTING LOG

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
7-7-11	Simulator	MI 3351900	(1) .082	G0101
		Certification Number 10636	(2) —	Signature D Dahlke #397
7-15-11	Simulator	MI 3351900	(1) .083	G0101
		Certification Number 10636	(2) —	Signature D. Dahlke #397
7-20-11	Simulator	MI 3351900	(1) .076	G0101
		Certification Number 19733	(2)	Signature #95
7-28-11	Simulator	MI 3351900	(1) .077	G0101
		Certification Number 19733	(2)	Signature #95
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	(1)	Ampoule Control Number
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature: D. Dahlke #397

Date: 8-29-11

OD-33 (9-2003)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	Year
AUGUST	2011

Instrument Location: LANSING POLICE

Instrument Number	ORI Number
931024	MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
8/3	SIMULATOR	MI 3351900	(1) .077	G0101
		Certification Number 11419	(2)	Signature Donald P[illegible] #147
8/9	SIMULATOR	MI 3351900	(1) .077	G0101
		Certification Number 11419	(2)	Signature Donald P[illegible] #147
8-13-11	Simulator	MI 3351900	(1) .082	G0101
		Certification Number 10636	(2) —	Signature D Dahlke #397
18	I have inspected, serviced, and checked this instrument for accuracy and certify that it is in proper working order Marvin Gler 120 Day Instrument Manufacturer's Authorized Representative	MI	(1) .079	J0701 Guth.080 [illegible] 12-2-11
		Certification Number 99998	(2) —	Signature Marvin Gler
8-21-11	Simulator	MI 3351900	(1) .082	J0701
		Certification Number 10636	(2) —	Signature D. Dahlke #397
8-29-11	Simulator	MI 3351900	(1) .082	J0701
		Certification Number 10636	(2) —	Signature D. Dahlke #397
		MI	(1)	
		Certification Number	(2)	Signature
		MI	(1)	
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature: D. Dahlke #397

Date: 9-7-11

OD-33 (9-2003)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	Year
SEPTEMBER	2011

Instrument Location: LANSING POLICE

Instrument Number	ORI Number
931024	MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
9-2-11	Simulator	MI 3351900	(1) .081	J 0701
		Certification Number 10636	(2) —	Signature D. Dahlke #397
9-13-11	Simulator	MI 3351900	(1) .080	J0701
		Certification Number 10636	(2) —	Signature D. Dahlke #397
9-19-11	Simulator	MI 3351900	(1) .082	J0701
		Certification Number 10636	(2) —	Signature D. Dahlke #397
9-26-11	Simulator	MI 3351900	(1) .081	J0701
		Certification Number 10636	(2) —	Signature D. Dahlke #397
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature
Day	Operation	Operator ORI Number MI	Results (1)	Ampoule Control Number
		Certification Number	(2)	Signature

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature	Date
D. W. Dahlke #397	10-2-11

OD-33 (9-2003)
MICHIGAN STATE POLICE
SPECIAL OPERATIONS DIVISION

AUTHORITY: 1949 PA 300
COMPLIANCE: Required
PENALTY: Removal/Decertification of Instrument

EVIDENTIAL BREATH TESTING LOG

Month	Year
OCTOBER	2011

Instrument Location: LANSING POLICE

Instrument Number: 931024

ORI Number: MI 3351900

Day	Operation	Operator ORI Number	Results	Ampoule Control Number
10-2-11	Simulator	MI 3351900	(1) .082	J0701
		Certification Number: 10636	(2) —	Signature: D. Dahlke #397
10-11-11	Simulator	MI 3351900	(1) .082	J0701
		Certification Number: 10636	(2) —	Signature: D. Dahlke #397
17	Operation: I have inspected, serviced, and checked this instrument for accuracy and certify that it is in proper working order. Marvin Geer 120 Instrument Manufacturer's Authorized Representative 0 of	MI	(1) .079	Ampoule Control Number: J-0701 Guth.080 to 09340 exp 12-2-2011
		Certification Number: 99998	(2) —	Signature: Marvin Geer
10-22-11	Simulator	MI 3351900	(1) .080	J0701
		Certification Number: 10636	(2) —	Signature: D. Dahlke #397
10/28/11	Simulator	MI 3351900	(1) .080	J0701
		Certification Number: 22096	(2)	Signature: B. Kelly 03
		MI	(1)	
			(2)	
		MI	(1)	
			(2)	
		MI	(1)	
			(2)	

Accurate test results shall be .076 to .084 inclusive.

Supervisor Signature

Date